

# HAFFKINE BIO PHARMACEUTICAL CORPORATION LIMITED Procurement Cell

( A Government of Maharashtra Undertaking)

Regd. Office: Acharya Donde Marg, Parel, Mumbai 400 012 (INDIA)

Phone No: 022- 24129320-23

Managing Director: 022-24150628

General Manager (Procurement Cell):022-

24100478

१. दिः २४.०७.२०१८ प्रशासकीय मंजूर निधी Approx.

Rate- 1,80,000/- (State Plan २०१८-१९) (Qty.- 19)

Website: http:/www.vaccinehaffkine.com

E-mail: procurementcell@haffkinemumbai.com

No.: 7582 /Haffkine/Procurement Cell/E-1093/

Operation Table (Hydraulic)/2021-22.

Date: 03 10 22

To,

M/s. Warden Surgical Co. Pvt. Ltd.,

75/76, Jawahar Co-operative Industrial Estate,

Kamothe, Panvel. Dist. Raigad - 410209.

Contact No.: 9324294452.

E-Mail: sales@wardensurgical.com

Subject : Supply Order for Tender No. E-1093/Operation Table (Hydraulic).

Reference: 1. Tender No. E-1093/HBPCL/PC/Operation Table (Hydraulic)/2018-19.

2. शासननिर्णय, क्रमांक - प्रशामा-२०१८/प्र.क.१६४/आरोग्य-८

दिनांक : २४ जुलै, २०१८.

3. Sanction of Tender Approval Committee Meeting No. 136

Dated :- 30.12.2021

With reference to the tender cited under reference no 1, you are requested to supply the following goods as per details mentioned below to consignee list enclosed with this order.

Sr. No.	Name of the item	Specification of item	Quantity/ Unit (DHS)	Unit Rate inclusive of GST(Rs.)	Total Amount Rs.
1	Operation Table (Hydraulic) Make : Warden Model : WHE-25	As per Annexure X	19	1,11,301/-	21,14,719/-

Total amount in words: Rupees Twenty One Lakh Fourteen Thousand Seven Hundred Nineteen Only.

Factory Location: 75/76, Jawahar Co-operative Industrial Estate, Kamothe, Panvel. Dist. Raigad - 410209.

- **Forwarding:** Forwarding Free on Road Destination. I.e. door delivery basis.
- 2 Delivery Period: 06 weeks from the date of receipt of order by the supplier to the consignee attached.
- Pre-Dispatch Inspection: Supplier shall make necessary arrangement / facilitate to carry out Pre-Dispatch inspection as per Tender Terms & condition and submit the Inspection report to this office. The Pre-Dispatch inspection cost will be borne by supplier. Machine should be dispatched only after Satisfactory Pre-Dispatch Inspection.

- Risk purchase clause: If the bidder fails to supply the stores within the stipulated delivery period, the order will stand cancelled. Undersigned shall be entitled to purchase such stores from any other source at such price which ordinarily should not be more than 10% of the tender price. The extra expenditure in such cases shall be recovered by Managing Director, Haffkine Bio Pharmaceutical Corporation Ltd. (Procurement Cell), Mumbai from the Supplier.
- Payment Terms: Payment of 100% of the contract value will be made within 8 weeks on delivery and successful installation and satisfactory commissioning and operation of the machinery.
- 6 Acceptance & Receipt: It should be submitted in Appropriate Format to the purchasing authority.
- 7 Delivery Challan Should be sent in the name of consignee in duplicate. It should specify Name of Equipment / Mfg. by / packing & quantity.
- 8 Invoice Copy Should be sent in triplicate on the Name of Managing Director, Haffkine Bio Pharmaceutical Corporation Ltd.(Procurement Cell), Mumbai. Along with Bill of Entry and Country of Origin Certificate of the consignment.

#### 9 Other Terms :

- 1) Warranty: The warranty period shall be for 2 years from the date of commissioning of all equipment supplied as certified by the consignee. After completion of 2 years warranty period Manufacturer/Supplier should give commitment to ensure services and supply of spare part for further 8 years. The successful tenderer must ensure 95% uptime during warranty period. In case of downtime, warranty period will be extended for period of downtime. If the equipment is not attended within 24 hours for Mumbai and 48 hours for other places the supplier will be liable to pay a penalty of 0.07% of purchase cost for every day of delay. Such penalty will be recovered from the amount of security deposit. Certificate of such uptime / downtime issued by the end user will be binding for the supplier Replacement of spares parts thereof due to manufacturing defects during warranty period will be entirely at the supplier's cost.
- 2) The user institution will enter to the Comprehensive Maintenance Contract with supplier agency @ 5% of the order value (excluding taxes) of the equipment per year for 8 years after completion of warranty period. In case of non-compliance of CMC the supplier will be liable to pay penalty or for appropriate action. Payment of CMC on yearly basis will be made by the user's institution, at the end of the year after satisfactory performance report from the end user.
- 10 Contract Agreement: Bidder should submit Contract Agreement on non-judicial stamp paper of requisite value.

#### **Fall Clause**

It is a condition of the contract that all through the currency there of, the price at which you will the supply stores should not exceed the lowest price charged by you to any customer during the currency of the contract and that in the event of the prices going down below the rate contract prices you shall promptly furnish such information to us to enable to amend the contract rates for subsequent supplies.

The Bidder should submit (within 7 days) amount of 1.5% i-e. Rs. 31,721/- of order value to meet other incidental expenditure and 3% i-e. Rs. 63,442/- as Security Deposit in form of Bank Guarantee. The Bank Guarantee valid for 2 months after the expiry of date of warranty issued by any Nationalized / Scheduled Bank.

## Amount to be deposited to Following Account:

Name of Account	Haffkine B P C L (Procurement Cell), CESS Account
Name of the Bank & Branch	Bank of Maharashtra, Branch- Mumbai Parel
Account No.	60381379835
IFSC Code	MAHB0000079

Consignee: As per list enclosed

मा. व्यवस्थापकीय संचालक बांच्या मान्यतेने व करिता

Dr. Sadanand Bhise General Manager Haffkine Bio Pharmaceutical Corporation Ltd. (Procurement Cell), Mumbai.

### Copy to:

- 1) Commissioner Health Services, Mumbai.
- 2) Director, Medical Education & Research, Mumbai-400 001.
- Account Manager, Haffkine Bio Pharmaceutical Corporation Ltd. (Procurement Cell), Mumbai.
- 4) Office File.

<u>Copy to Consignee:</u> As Per Consignee List.: As per Tender Condition No.17 The user Institution should get the Comprehensive Maintenance Contract done with supplier agency @ 5% of the Order value (excluding taxes) of equipment per year for Eight years after Completion of warranty period.

Copy Submitted to: 1) Secretary, Medical Education & Drug Department, Mantralaya, Mumbai.

## Annexure-X

	nical Specification for Operation Table (Hydraulic)
GMDN Name	Operation Table Hydraulic Major
Clinical purpose	An operation table, sometimes called operating room table, is the table
	on which the patient lies during a surgical operation, This surgical
	equipment is usually found inside the surgery room of a hospital.
Used by clinical	Operation theatre.
Department/ward	
TECHNICAL CHARACT	ERISTICS (SPECIFIC TO THIS TYPE OF DEVICE)
	lly controlled operating table, working range from floor level: 800-
1040mm ± 10%, by	
	le to all essential positions.
	d with movement controls at side of the table.
<ul> <li>Should have Frame</li> </ul>	e and bottom Made of Stainless Steel 304 material
<ul> <li>Should have reinfo</li> </ul>	rced three section stainless steel top.
<ul> <li>Height should be a</li> </ul>	djustable by oil pump from $800\text{mm}-1040\text{mm} \pm 10\%$ , by foot step control.
<ul> <li>Should have detach</li> </ul>	nable head rest which can be easily adjustable to any desired position,
above or below tab	
	stated 360° through base.
<ul> <li>Trendelenburg : ≥ 3</li> </ul>	25°-30°
<ul> <li>Reversed Trendele</li> </ul>	nburg: ≥30°
Head Section Raise	ed from the Horizontal: ≥20°-30°
Head Section Lower	ered from the Horizontal: ≥28°-30°
	ed from the Horizontal: ≥60°-70°
	red from the Horizontal: ≥40°-50°
	ould be achievable by breaking the table.
Table-top should b	
User's interface Setting	Manual
Dimensions (metric)	Table top dimension (1900 mm x 525 mm) ± 15%
Difficusions (metric)	Table elevation: $800 \text{ mm} + 1040 \text{ mm} \pm 10\%$
Weight (lbs, kg)	Should be able to bear patient having weight up to 180 kg
Heat dissipation	Heat Dissipation: Should maintain nominal Temp and the heat should be
A	disbursed through a cooling mechanism.  1. S. S. Arm Rest 1 No
Accessories (mandatory)	Color 19 - Depter and the control of the color of the col
	2. Anesthetic Screen 1 No.
	3. Lithotomy Leg Holders with stir- Ups 1 Set
	4. Leather Wristlets 1 Set
	5. Padded Leg Rest (Gutter Type) – 2 Nos
	6. Anti-static mattress – 2 Nos
	7. Side rails- 2 Nos
Atmosphere / Ambiance	1) Operating condition: Capable of operating continuously in ambient
(air conditioning,	temperature of 10 to 40deg C and relative humidity of 15 to 90% in idea
humidity, dust)	circumstances.
	2) Storage condition: Capable of being stored continuously in ambient
	temperature of 0 to 50 deg C and relative humidity of 15 to 90%.
User's care, Cleaning,	1) Disinfection: Parts of the Device that are designed to come into
Disinfection & Sterility	contact with the patient or the operator should either be capable of easy
issues	disinfection or be protected by a single use/disposable cover.
	2) Sterilization not required
Certificates	1) Should be FDA/CE/BIS approved product
Local and / or	Manufacturer/supplier should have ISO certification for quality standard

international	
Pre-installation	Safety and operation check before handover
requirements nature,	
values, quality, tolerance	
Requirements for sign-	Certification of calibration and inspection from the manufacturer
off	
Training of staff	Training of users in operation and basic maintenance shall be provided
(medical, paramedical,	Advanced maintenance tasks required shall be documented
technicians)	•
Warranty	Three years on site with free servicing (min.3 year during warranty &
	provide technical support and required spares and consumable for 7 yrs
	after warranty period.
Maintenance tasks	Maintenance manual detailing complete maintaining schedule
Service contract clauses,	The spare price list of all spares and accessories (including minor)
including prices	required for maintenance and repairs in future after guarantee / warranty
	period should be attached.
Operating manuals,	1) User, technical and maintenance manuals to be supplied in
service manuals, other	English/Marathi Hindi language along with machine diagrams.
manuals	List of equipment and procedures required for local calibration and
and the second control of the second control	routine maintenance.
	3) Service and operation manuals (original and copy) to be provided.
	4) Advanced maintenance tasks documentation.
	5)Leather Wristlets 1 Set
	6) Padded Leg Rest (Gutter Type) – 2 Nos
	7) Anti-static mattress – 2 Nos
	8) Side rails – 2 Nos
Atmosphere / Ambiance	Operating condition: Capable of operating continuously in ambient
(air conditioning,	temperature of 10 to 40deg C and relative humidity of 15 to 90% in ideal
humidity, dust)	circumstances.
indifficity, dust)	2) Storage condition: Capable of being stored continuously in ambient
User's care, Cleaning,	temperature of 0 to 50 deg C and relative humidity of 15 to 90%.  1) Disinfection: Parts of the Device that are designed to come into
Disinfection & Sterility	contact with the patient or the operator should either be capable of easy
issues	disinfection or be protected by a single use/disposable cover.
133463	2) Sterilization not required
Certificates	Should be FDA/CE/BIS approved product.
Local and/or	Manufacturer/supplier should have ISO certificate for quality standard
international	Walturacturer/supplier should have 150 certificate for quality standard
Pre-installation	Safety and operation check before handover
requirements nature,	Safety and operation check before nandover
values, quality, tolerance	9
Requirements for sign-	Cartificate of calibration and inspection from the manufacture
off	Certificate of calibration and inspection from the manufacturer
Training of staff	Training of users in operation and basis maintenance shall be and it is
(medical, paramedical,	Training of users in operation and basic maintenance shall be provided
technicians)	Advanced maintenance tasks required shall be documented
Warranty	Three years on site with free comicing (min 2
" arranty	Three years on site with free servicing (min 3 year during warranty &
	provide technical support and required spares and consumable for 7 yrs
Maintenance tasks	after warranty period.
	Maintenance manual detailing complete maintaining schedule
Service contract clauses,	The spare price list of all spares and accessories (including minor)
including prices	required for maintenance and repairs in future after guarantee / warranty
EST to least beller toma a very consent.	period should be attached.

Operating manuals, service manuals, other manuals	1) User, technical and maintenance manuals to be supplied in English/Marathi Hindi language along with machine diagrams.
	2) List of equipment and procedures required for local calibration and routine maintenance.
* ************************************	3) Service and operation manuals (original and copy) to be provided.
	4) Advanced maintenance tasks documentation.
	5) Certificate of calibration and inspection
Other accompanying documents	List to be provided of important spares and accessories, with their part numbers and cost,
Service, Support Contact details (Hierarchy Wise; including a toll free/landline number)	Contact details of manufacturer, supplier and local service agent to be provided;
Recommendations or warnings	Any warning signs would be adequately displayed

# **Consignee Details**

	Operation Table (Hydraulic) Make: Warden & Model: WHE-25			
Delivery Period	06 weeks  No.: 7582 /Haffkine/ Procurement Cell/ E-1093/ Operation Table (Hydraulic)/ 2022-23.  Date: 03   10   22			
PO Ref. No.				
दिः २४.०७.२०	१८ प्रशासकीय मंजूर निधी Approx. Rate- 1,80,000/- (State Plan २०१८	-१९) (Qty 19)		
Sr. No.	Name & Address of the Consignee	Qty.		
1)	Rural Hospital, Jeur, Solapur.	01		
2)	Rural Hospital, Ashti, Wardha.	01		
3)	Sub District Hospital, Karanja, Washim.	02		
4)	Rural Hospital, Sonpeth, Parbhani.	01		
5)	Rural Hospital, Bhusawal, Jalgaon.	01		
6)	Rural Hospital, Sawada, Jalgaon.	01		
7)	Rural Hospital, Aurad Shahajani, Latur.	01		
8)	Rural Hospital, Shirur Anantpal, Latur.	01		
9)	Women Hospital, Malegaon, Nashik.	02		
10)	Women Hospital, Dhule.	02		
11)	Women Hospital, Ambajogai, Beed.	02		
12)	Women Hospital, Washim.	02		
13)	Women Hospital, Buldhana.	02		
	Total	19		

मा. व्यवस्थापकीय संचालक यांच्या मान्यतेने व करिता

Dr. Sadanand Bhise General Manager

Haffkine Bio Pharmaceutical Corporation Ltd. (Procurement Cell), Mumbai.