

**HAFFKINE BIO PHARMACEUTICAL CORPORATION LIMITED****Procurement Cell****(A Government of Maharashtra Undertaking)****Regd. Office : Acharya Donde Marg, Parel, Mumbai 400 012 (INDIA)****Phone No: 022- 24129320-23****Website: <http://www.vaccinehaffkine.com>****Managing Director :022-24150628****E-mail: procurementcell@vaccinehaffkine.com****Corrigendum.****No.: 1874 / E-2775/ Change in Technical Specification / HBPC/ PC/ 2021-22.****Date: 10/02/2022.****To,****M/s. Bunty Furniture & Rehabilitation Aids.,****Behind Chandralock Building,****Near IGM Hospital, Old Bhandara Road,****Nagpur – 440018.****Email ID: bunty.777@rediffmail.com****Subject : - Change in Technical Specification****Reference: - 1.Tender no. E-2775/Dressing Trolley****2. Purchase order No.6321-6010/Haffkine /Procurement Cell/ E-2775****Dated: 29/12/2021-04/02/2022 .****As per above mentioned / Change in Technical Specification**

Read As	Instead of
Overall Size :- 1210 mm L x 510 mm W x Ht 870 mm (Approx.)/48"X 34" ht.	Overall Size :- 760 mm L x 510 mm W x Ht 870 mm (Approx.)/30" x 20" X 34" ht.
Complete SS 394 Grade.	All vertical sections made form 25.4 x 25.4mm x 18 swg S.S tube.
All vertical sections made form 25.4 x 25.4mm x 18 swg S.S tube.	
Top and bottom should be made from 1 mm thick S.S sheet double press bent on all 4 sides	Top and bottom should be made from 1.mm thick S.S sheet double press bent on all 4 sides
Provided with 100 mm twin wheel, non-rusting & non marking revolving castors made form high strength engineering plastic.	Provided with 100 mm twin wheel, nonrusting & non marking revolving castors made form high strength engineering plastic.
The Trolley Should be Provided with all 4 locking Castors	All the four wheels should be provided with locks.
Railings made from 12mm dia SS tube 18 swg tube to be provided on three sides of the shelf and the top	Railings made from 12mm dia S.s tube 18 swg tube to be provided on three sides of the shelf and the top
Top end of vertical tube to be closed with S.S. sheet with seamless joint.	Top end of vertical tube to be closed with S.S. sheet with seamless joint.
Handle to be provided on end of the trolley in 19 mm dia S.S tube.	The Trolley should be provided with 100 mm deep drawer on high quality telescopic channels.
Top and shelves is matt finish to prevent slippage hands and items kept on shelves and resist form scratches.	The drawer should be provided with Self-clinching nuts placed in properly sized holes in the sheet and then fitted by applying a parallel squeezed force on the head. Which make the self-clinching nut an integral part of the sheet drawer. These self-clinching nuts are provided as to ease the fitting of the telescopic channel on to the drawer.

Trolley to be provided with detachable ring in 16mm S.S 304 grade tube.	The other side of the telescopic Chanel to be secured to the internal side of the drawer in the same fashion as it has been fixed to the pull out drawer.
The bowl to be 15" dia x 4" deep.	The drawer should have built in pull out handles in press bent S.S sheet itself. No external handles to be provided that protrude out of the trolley.
Bucket to be 9.5 dia and 8" deep. Bucket should be provided with lid.	Top and shelves is matt finish to prevent slippage hands and items kept on shelves and resist form scratches.
Completely seamless weld at joints	Completely seamless weld at joints
Load bearing capacity should be at least 70 kg.	Load bearing capacity should be at least 60 kg.
Complete trolley should be supplied as a single unit and not in knock down condition.	Complete trolley should be supplied as a single unit and not in knock down condition.

All other Terms & Conditions of the order will remain unchanged.

मा. व्यवस्थापकीय संचालक यांच्या मान्यतेने व करिता

(General Manager)

Haffkine Bio-Pharmaceutical Corporation Ltd,
(Procurement Cell), Mumbai- 1