

**HAFFKINE BIO PHARMACEUTICAL CORPORATION LIMITED****Procurement Cell****( A Government of Maharashtra Undertaking)****Regd. Office : Acharya Donde Marg, Parel, Mumbai 400 012 (INDIA)****Phone No: 022- 24129320-23****Managing Director :022-24150628****General Manager (Procurement Cell) :022-24100478****Website : http://www.vaccinehaffkine.com****E-mail: procurementcell@vaccinehaffkine.com****No.: 4216 /Haffkine/Procurement Cell/E-2348/  
Operation Table (Hydraulic)/ 2020-21.****Date: 29/11/2021****To,****M/s. Galaxy India Pvt. Ltd.,****T-5/237, Mangolpuri Indl. Area,****Phase-I, New Delhi - 110083.****Contact No. 9999926655.****E-Mail: galaxyindia.delhi@gmail.com****Subject : Supply Order for Tender No. E-2348/Operation Table (Hydraulic)****Reference: 1. Tender No. E-2348/HBPCL/PC/ Operation Table (Hydraulic)/ 2019-20.****2. Sanction of Tender Approval Committee Meeting Dated 23.12.2020****3. प्रशामा- खरेदी-२०१९/प्र.क्र.१०८/आरोग्य-८ दिनांक: १८ सप्टेंबर, २०१९.**

With reference to the tender cited under reference no 1, you are requested to supply the following goods as per details mentioned below to consignee list enclosed with this order.

Sr. No.	Name of the item	Specification of item	Quantity/ Unit (DHS)	Unit Rate inclusive of GST(Rs.)	Total Amount Rs.
1	<b>Operation Table (Hydraulic) Make: M/s. Galaxy India Model: Relax</b>	As per Annexure X	32	1,71,808/-	<b>54,97,856/-</b>

**Total amount in words: Rupees Fifty Four Lakhs Ninety Seven Thousand Eight Hundred Fifty Six Only.****Factory Location: T-5/237, Mangolpuri Indl. Area, Phase-I, New Delhi - 110083.**

- Forwarding:** Forwarding Free on Road Destination. I.e. door delivery basis.
- Delivery Period:** 06 weeks from the date of receipt of order by the supplier to the consignee attached.
- Pre-Dispatch Inspection:** Supplier shall make necessary arrangement / facilitate to carry out Pre-Dispatch inspection as per Tender Terms & condition and submit the Inspection report to this office. The Pre-Dispatch inspection cost will be borne by supplier. Machine should be dispatched only after satisfactory Pre-Dispatch Inspection.

- 12 The Bidder should submit (within 7 days) amount of 1.5% ie. **Rs. 82,468/-** of order value to meet other incidental expenditure and 3% i-e **Rs. 1,64,936/-** as Security Deposit in form of Bank Guarantee. The Bank Guarantee valid for 2 months after the expiry of date of warranty issued by any Nationalized / Scheduled Bank.

**Amount to be deposited to Following Account:**

Name of Account	Haffkine Bio-Pharmaceutical Corporation Ltd. (Procurement Cell), CESS Account Mumbai.
Name of the Bank & Branch	Canara Bank, Branch-Parel
Account No.	0110201004893
IFSC Code	CNRB0000110

**Consignee:** As per list enclosed

मा. व्यवस्थापकीय संचालक यांच्या मान्य टिपणी नुसार



(Dr. R. M. Kumbhar)

General Manager - 1

**Haffkine Bio Pharmaceutical Corporation Ltd.**  
(Procurement Cell), Mumbai.

**Copy to:**

- 1) Commissioner Health Services, Mumbai.
- 2) Director, Medical Education & Research, Mumbai-400 001.
- 3) Account Manager, Haffkine Bio Pharmaceutical Corporation Ltd. (Procurement Cell), Mumbai.
- 4) Office File.

**Copy to Consignee:** As per Consignee list. : As per Tender Condition No.17 The user Institution should get the Comprehensive Maintenance Contract done with supplier agency @ 5% of the Order value (excluding taxes) of equipment per year for Eight years after Completion of warranty period.

**Copy Submitted to:** 1) Secretary, Medical Education & Drug Department, Mantralaya, Mumbai.

## Annexure-X

Operation Table (Hydraulic)

1.	Clinical Purpose	An operating table, sometimes called operating room table, is the table on which the patient lies during a surgical operation. This surgical equipment is usually found inside the surgery room of a hospital.
2.	Used by clinical department / ward	Operating theatre
3.	Technical characteristics specific to this type of devise	<ol style="list-style-type: none"> <li>1) Should be a manually controlled operating table, working range from floor level: 800-1040 mm <math>\pm</math> 10%</li> <li>2) Should be adjustable to all essential positions.</li> <li>3) Should be equipped with movement controls at side of the table.</li> <li>4) Should have frame and bottom made of stainless steel 304 material</li> <li>5) Should have reinforced three section stainless steel top.</li> <li>6) Height should be adjustable by oil pump from 28" to 42", by foot step control.</li> <li>7) Should have detachable head rest which can be easily adjustable to any desired position, above or below table top.</li> <li>8) Table top can be rotated 360° through base.</li> <li>9) Trendelenburg : 25° - 30°</li> <li>10) Reversed trendelenburg : ..30°</li> <li>11) Head section Raised from the Horizontal : 20°-30°</li> <li>12) Head section Lowered from the Horizontal: 28° - 30°</li> <li>13) Back section Raised from the Horizontal: &gt;60°- 70°</li> <li>14) Leg section Lowered from the Horizontal: 40°-50°</li> <li>15) Kidney position should be achievable by breaking the table.</li> <li>16) Table-top should be radio-lucent.</li> </ol>
4.	User's interface	Manual
5.	Dimensions (metric)	Table top dimension (1900mm x 525 mm) $\pm$ 15% table elevation: (700mm – 1000mm) $\pm$ 10%
6.	Weight (lbs, kg)	Should be able to bear patient having weight up to 180 kg
7.	Heat dissipation	Heat dissipation: should maintain Temp and the heat should be disbursed through an cooling mechanism
8.	Accessories (mandatory)	<ol style="list-style-type: none"> <li>1) S. S. Arm Rest 1 No</li> <li>2) Anesthetic screen 1 No.</li> <li>3) Lithotomy leg holders with stir-Ups 1 set</li> <li>4) Leather wristlets 1 Set</li> <li>5) Padded leg Rest (Gutter Type)- 2 Nos</li> <li>6) Anti-static matterss- 2 Nos</li> <li>7) Side rails- 2 Nos</li> </ol>
9.	Atmosphere/Ambiance (air conditioning, humidity, dust..)	1) Operating condition: Capable of operating continuously in ambient temperature of 10 to 40 deg C and relative humidity of 15 to 90% in ideal

		<p>circumstances.</p> <p>2) Storage condition: Capable of being stored continuously in ambient temperature of 0 to 50 dec C and relative humidity of 15 to 90%</p>
10.	User's care, Cleaning Disinfection & Sterility issues	<p>1) Disinfection: Parts of the Device that are designed to come into contact with the operator should either be capable of easy disinfection or be protected by a single use/ disposable cover.</p> <p>2) Sterilization not required.</p>
11.	Certificates	Should be USFDA registered.
12.	Local and/ or international	Manufacture/supplier should have ISO certificates for quality standard.
13.	Pre-installation requirement: Nature, Values, Quality, tolerance	Safety and operation check before handover.
14.	Requirement for sign-off	Certificate of calibration and inspection form the manufa
15.	Training of staff (Medical, paramedical, technicians)	<p>1) Training of users on operation and basic maintenance;</p> <p>2) Advanced maintenance tasks required shall be documented</p>
16.	Warranty	Two year on site with free servicing (min. 03/year) during warranty & provided technical support and required spares and consumable for 7 yrs after warranty period.
17.	Maintenance task	<p>1) Maintenance manual detailing.</p> <p>2) Complete maintenance schedule.</p>
18.	Service contract clauses, including prices	The spare price list of all spare and accessories (including minor) required for maintenance and repair in future after guarantee warranty period should be attached.
19.	Operating manuals, service manuals, other manuals	<p>Should provide 2 sets (hardcopy and soft-copy) of:-</p> <p>1) User's technical and maintenance manuals to be supplied in English/Marathi/Hindi language along with machine diagrams.</p> <p>2) List of equipment and procedures required for local calibration and routine maintenance.</p> <p>3) Service and operation manuals (original and copy) to be provided.</p> <p>4) Advanced maintenance tasks documentation.</p> <p>5) Certificate of calibration and inspection.</p>
20.	Other accompanying documents	List of important spares and accessories, with their part numbers and cost.
21.	Service support contact details (Hierarchy Wise; including a toll free/landing number)	Contact details to manufacturer, supplier and local service agent to be provided;
22.	Recommendation or warning	Any warning sings would be adequately displayed.

### Consignee Details

M/s. Galaxy India Pvt. Ltd.		
Operation Table (Hydraulic) Make: M/s. Galaxy India Model: Relax		
Delivery Period	06 weeks	
PO Reference No.	No.: 4216 /Haffkine/Procurement Cell/E-2348/Operation Table (Hydraulic)/2020-21. Date: 29/11/2021	
Sr. No.	Name & Address of the Consignee	Qty.
1)	Rural Hospital, Alandi, Pune.	1
2)	Rural Hospital, Taleghar, Pune.	1
3)	Rural Hospital, Pombhurna, Chandrapur.	1
4)	Rural Hospital, Bhatkuli, Amarawati.	1
5)	Rural Hospital, Rajur, Jalna.	1
6)	Rural Hospital, Kokrud, Sangli.	1
7)	Rural Hospital, Hivra Ashram, Buldhana.	1
8)	Sub District Hospital, Warora, Chandrapur.	2
9)	Sub District Hospital, Indapur, Pune.	1
10)	Sub District Hospital, Rajur, Chandrapur.	2
11)	Sub District Hospital, Arvi, Wardha.	1
12)	District Hospital, Buldhana.	4
13)	Women Hospital, Ahamednagar.	2
14)	Women Hospital, Jalna.	3
15)	Women Hospital, Sindhudurgh.	2
16)	Women Hospital, Achalpur, Amarawati.	2
17)	General Hospital, Khamgaon, Buldhana.	2
18)	Indira Gandhi Memorial Hospital, Ichalkarnaji, Kolhapur.	4
	<b>Total</b>	<b>32</b>

मा. व्यवस्थापकीय संचालक यांच्या मान्य दिपणी नुसार



(Dr. R. M. Kumbhar)

General Manager - 1

Haffkine Bio Pharmaceutical Corporation Ltd.  
(Procurement Cell), Mumbai.