Haffkine HAFFKINE BIO PHARMACEUTICAL CORPORATION LIMITED **Procurement Cell** (A Government of Maharashtra Undertaking) Regd. Office : Acharya Donde Marg, Parel, Mumbai 400 012 (INDIA) Phone No: 022- 24129320-23 Website : http://www.vaccinehaffkine.com Managing Director :022-24150628 E-mail: procurementcell@vaccinehaffkine.com General Manager (Procurement Cell) :022-24100478 No.: 7746 /Haffkine/Procurement Cell/E-4444/Autoclave प्रशासकीय मान्यता निधी २४,००,०००/-Machine/2022-23. राज्य योजना २०२१-२२ Date: 15/10/22

To,

4 64

M/s. Equitron Medica Private Limited

1st Floor, Unit No. 116,
Vasan Udyog Bhavan,
Sun Mill Compound,
S B Marg, Lower Parel West,
Mumbai - 400 013.
E-Mail: <u>bhushans@medicainstrument.com</u>

Subject : Supply Order for Tender No. E-4444/Autoclave Machine.

Reference: 1. Tender No. E-4444/HBPCL/PC/Autoclave Machine/2021-22.

2. शा.नि.क्र. : सीएचएम-२०२१/प्र.क्र.२६२/ प्रशा-१ दि. २० सप्टेंबर, २०२१.

3. Sanction of Tender Approval Committee Meeting No. 163 Dated 12.08.2022.

With reference to the tender cited under reference no 1, you are requested to supply the following goods as per details mentioned below to consignee list enclosed with this order.

Sr. No.	Name of the item	Specification of item	Quantity/ Unit (DMER)	Unit Rate inclusive of GST(Rs.)	Total Amount Rs.
1.	Autoclave Machine Make: Equitron Model : #7410-4072	As per Annexure X	04	4,95,600/-	19,82,400/-
Total	amount in words: Rupees Nir	neteen Lakhs Eighty	Two Thousand	d Four Hundred On	ıly.
Factor	ry Location: EL-177, Electron	ic Zone, TTC Indus	strial Area, MII	DC, Mahape – 4007	10.

1 **Forwarding:** Forwarding Free on Road Destination. I.e. door delivery basis.

2 **Delivery Period**: 06 weeks from the date of receipt of order by the supplier to the consignee attached.

3 Pre-Dispatch Inspection: Supplier shall make necessary arrangement / facilitate to carry out Pre-Dispatch inspection as per Tender Terms & condition and submit the Inspection report to this office. The Pre-Dispatch inspection cost will be borne by supplier. Machine should be dispatched only after satisfactory Pre-Dispatch Inspection.

- 4 Risk purchase clause: If the bidder fails to supply the stores within the stipulated delivery period, the order will stand cancelled. Undersigned shall be entitled to purchase such stores from any other source at such price which ordinarily should not be more than 10% of the tender price. The extra expenditure in such cases shall be recovered by Managing Director, Haffkine Bio Pharmaceutical Corporation Ltd. (Procurement Cell), Mumbai from the Supplier.
- 5 **Payment Terms:** Payment of 100% of the contract value will be made within 8 weeks on delivery and successful installation and satisfactory commissioning and operation of the machinery.
- 6 Acceptance & Receipt: It should be submitted in Appropriate Format to the purchasing authority.
- 7 **Delivery Challan** Should be sent in the name of consignee in duplicate. It should specify Name of Equipment / Mfg. by / packing & quantity.
- 8 **Invoice Copy** Should be sent in triplicate on the Name of Managing Director, Haffkine Bio Pharmaceutical Corporation Ltd.(Procurement Cell), Mumbai. Along with Bill of Entry and Country of Origin Certificate of the consignment.

9 Other Terms :

1) Warranty: The warranty period shall be for 2 years from the date of commissioning of all equipment supplied as certified by the consignee. After completion of 2 years warranty period Manufacturer/Supplier should give commitment to ensure services and supply of spare part for further 8 years. The successful tenderer must ensure 95% uptime during warranty period. In case of downtime, warranty period will be extended for period of downtime. If the equipment is not attended within 24 hours for Mumbai and 48 hours for other places the supplier will be liable to pay a penalty of 0.07% of purchase cost for every day of delay. Such penalty will be recovered from the amount of security deposit. Certificate of such uptime / downtime issued by the end user will be binding for the supplier Replacement of spares parts thereof due to manufacturing defects during warranty period will be entirely at the supplier's cost.

2) The user institution will enter to the Comprehensive Maintenance Contract with supplier agency @ 5% of the order value (excluding taxes) of the equipment per year for 8 years after completion of warranty period. In case of non-compliance of CMC the supplier will be liable to pay penalty or for appropriate action. Payment of CMC on yearly basis will be made by the user's institution, at the end of the year after satisfactory performance report from the end user.

10 **Contract Agreement:** Bidder should submit Contract Agreement on non-judicial stamp paper of requisite value.

Fall Clause

It is a condition of the contract that all through the currency thereof, the price at which you will the supply stores should not exceed the lowest price charged by you to any customer during the currency of the contract and that in the event of the prices going down below the rate contract prices you shall promptly furnish such information to us to enable to amend the contract rates for subsequent supplies.

Page 2 of 5

DOL

11 The Bidder should submit (within 7 days) amount of 1.5% i-e. **Rs. 29,736/-** of order value to meet other incidental expenditure and 3% i-e. **Rs. 59,472/-** as Security Deposit in form of Bank Guarantee. The Bank Guarantee valid for 2 months after the expiry of date of warranty issued by any Nationalized / Scheduled Bank.

Amount to be deposited to Following Account:

Name of Account	Haffkine B P C L (Procurement Cell), CESS Account	
Name of the Bank & Branch	Bank of Maharashtra, Branch- Mumbai Parel	
Account No.	60381379835	
IFSC Code	MAHB0000079	

Consignee: As per list enclosed

Dr. Sadanand Bhise General Manager Haffkine Bio Pharmaceutical Corporation Ltd. (Procurement Cell), Mumbai.

मा. व्यवस्थापुकीय संचौलक यांच्या मान्यतेने व करिता

Copy to:

- 1) Commissioner Health Services, Mumbai.
- 2) Director, Medical Education & Research, Mumbai-400 001.
- Account Manager, Haffkine Bio Pharmaceutical Corporation Ltd. (Procurement Cell), Mumbai.
- 4) Office File.
- 5) System Integrator.

<u>Copy to</u> Consignee: Dean, CAMA & Albless Hospital, Mumbai.: As per Tender Condition No.17 The user Institution should get the Comprehensive Maintenance Contract done with supplier agency @ 5% of the Order value (excluding taxes) of equipment per year for Eight years after Completion of warranty period.

Copy Submitted to: 1) Secretary, Medical Education & Drug Department, Mantralaya, Mumbai.

 Specifications for Autoclave Machine The Vertical top loading Autoclave should have Chamber Capacity of 80 to 90 ltr. / drying with chamber dimension (030 x 50 cm approx.). The Unit should be able to accommodate 01 carrier of size 25cm x 45cm. The unit's internal chamber, cover lid, external body and all other wetted parts should be fabricated from stainless steel of 304 grade. The Vertical Autoclave's all joints should be TIG Welded with inert gas and smooth finished for Crevice free internals. The autoclave should have hydraulic die press scamless lid from stainless locking s/m steel plate. The autoclave should have locking / system for lid. Should have insulated material (industrial plastic) – cool to touch type. The lid cover should be provided with auto purge cum vacuum breaker valve and a manually operable valve for exhaust. The unit should have a solenoid valve for additionally purging the air & end of cycle automatic exhaust. The Vertical Autoclave should have stainless steel pressure gauge with dual range dial display in KPA and PSI along with a correlated temperature scale for steam in degrees Celsius. The operations of the unit should be controlled by a microprocessor-based controller. The user should be able to set the temperature up to 122°C in steps of 0.1°C each. The user should be able to set the Sterilization hold time in steps of 1 minutes each. The timer range should be upto 95 minutes In case of abrupt power failure / switching off, the Last Cycle which was set should remain in the memory. The unit should have stafety valve to protect the equipment in case of orver pressurization. The unit should ba equipped with Low Water Detection unit and should give Audio- Visual alarm in case of Low Water in the chamber and cut off the supply to the heater. The unit should have stafety valve to protect the equipment in case of orver pressurizat		Annexure-X
 with chamber dimension (Ø30 x 50 cm approx.). The Unit should be able to accommodate 01 carrier of size 25cm x 45cm. The unit's internal chamber, cover lid, external body and all other wetted parts should be fabricated from stainless steel of 304 grade. The Vertical Autoclave's all joints should be TIG Welded with inert gas and smooth finished for Crevice free internals. The autoclave should have hydraulic die press seamless lid from stainless locking s/m steel plate. The autoclave lid should have locking / system for lid. Should have insulated material (industrial plastic) – cool to touch type. The lid cover should be provided with auto purge cum vacuum breaker valve and a manually operable valve for exhaust. The unit should have a solenoid valve for additionally purging the air & end of cycle automatic exhaust. The Vertical Autoclave should have stainless steel pressure gauge with dual range dial display in KPA and PSI along with a correlated temperature scale for steam in degrees Celsius. The operations of the unit should be controlled by a microprocessor-based controller. The user should be able to set the Sterilization hold time in steps of 1 minutes each. The time range should be upto 95 minutes In case of abrupt power failure / switching off, the Last Cycle which was set should remain in the memory. The unit should be equipped with Low Water Detection unit and should give Audio- Visual alarm in case of Low Water in the chamber and cut off the supply to the heater. (pdv) The unit should have safety valve to protect the equipment in case of over pressurization. The unit should have safety valve to protect the equipment in case of over pressurization. The unit should be spressure equipment & hence from safety point of view chamber should be vertical Autoclave should be CE(NB) certified. The unit should be suppled with Low Water Detection unit and s		Specifications for Autoclave Machine
 The Unit should be able to accommodate 01 carrier of size 25cm x 45cm. The unit's internal chamber, cover lid, external body and all other wetted parts should be fabricated from stainless steel of 304 grade. The Vertical Autoclave's all joints should be TIG Welded with inert gas and smooth finished for Crevice free internals. The autoclave should have hydraulic die press seamless lid from stainless locking s/m steel plate. The autoclave lid should have locking / system for lid. Should have insulated material (industrial plastic) – cool to touch type. The lid cover should be provided with auto purge cum vacuum breaker valve and a manually operable valve for exhaust. The vortical Autoclave should have a solenoid valve for additionally purging the air & end of cycle automatic exhaust. The Vertical Autoclave should have stainless steel pressure gauge with dual range dial display in KPA and PSI along with a correlated temperature scale for steam in degrees Celsius. The operations of the unit should be controlled by a microprocessor-based controller. The user should be able to set the Eterilization hold time in steps of 1 minutes each. The display (LED/LCD) for the parameters should be two line Alpha memoric digital display. The autoclave should give sensor open alarm in case sensor fault. The autoclave should be ensured by inbuilt MCB. The autoclave should be ensured by inbuilt MCB. The unit should be ensured by inbuilt MCB. The vertical Autoclave is pressure equipment in case of its working pressure. The vertical Autoclave is pressure equipment & ensor fault. The unit should be ensured by inbuilt MCB. The autoclave should by ensure by inbuilt MCB. The unit should be controlled by a micro first of its working pressure. The vertical Autoclave is pressure equipment & hence from safety point of view chamber should be designed as per	•	The Vertical top loading Autoclave should have Chamber Capacity of 80 to 90 ltr. / drying
 The unit's internal chamber, cover lid, external body and all other wetted parts should be fabricated from stainless steel of 304 grade. The Vertical Autoclave's all joints should be TIG Welded with inert gas and smooth finished for Crevice free internals. The autoclave should have hydraulic die press seamless lid from stainless locking s/m steel plate. The autoclave lid should have locking / system for lid. Should have insulated material (industrial plastic) – cool to touch type. The lid cover should be provided with auto purge cum vacuum breaker valve and a manually operable valve for exhaust. The unit should have a solenoid valve for additionally purging the air & end of cycle automatic exhaust. The Vertical Autoclave should be controlled by a microprocessor-based controller. The operations of the unit should be controlled by a microprocessor-based controller. The user should be able to set the temperature up to 122°C in steps of 0.1°C each. The time range should be upto 95 minutes In case of abrupt power failure / switching off, the Last Cycle which was set should remain in the memory. The display (LED/LCD) for the parameters should be two line Alpha memoric digital display. The autoclave should be quipped with Low Water Detection unit and should give Audio- Visual alarm in case of Low Water in the chamber and cut off the supply to the heater. The autoclave should be upto vary to protect the equipment in case of over pressurization. The autoclave should be due to statication by a microprocestor fault. The autoclave should be due to set protect the equipment in case of over pressurization. The display (LED/LCD) for the parameters should be two line Alpha memoric digital display. The autoclave should be upto vary optoclet the equipment in case of over pressurization. The autoclave should be upto vary optoclet the equipment in case of over pr		
 fabricated from stainless steel of 304 grade. The Vertical Autoclave's all joints should be TIG Welded with inert gas and smooth finished for Crevice free internals. The autoclave should have hydraulic die press scamless lid from stainless locking s/m steel plate. The autoclave lid should have locking / system for lid. Should have insulated material (industrial plastic) – cool to touch type. The lid cover should be provided with auto purge cum vacuum breaker valve and a manually operable valve for exhaust. The unit should have a solenoid valve for additionally purging the air & end of cycle automatic exhaust. The Vertical Autoclave should have stainless steel pressure gauge with dual range dial display in KPA and PSI along with a correlated temperature scale for steam in degrees Celsius. The operations of the unit should be controlled by a microprocessor-based controller. The user should be able to set the temperature up to 122°C in steps of 0.1°C each. The timer range should be upto 95 minutes In case of abrupt power failure / switching off, the Last Cycle which was set should remain in the memory. The display (LED/LCD) for the parameters should be two line Alpha memoric digital display. The autoclave should give sensor open alarm in case of over pressurization. The autoclave should be valve to protect the equipment in case of over pressurization. The autoclave should be bydrostatically tested at 1.5 times of its working pressure. The Vertical Autoclave is pressure equipment & hence from safety point of view chamber should be CE(NB) certified. Autoclave should be SUB by CE(NB) certified. Manufacturer shall be ISO 13485 certified & should submit photocopy for the same. 	•	The Unit should be able to accommodate 01carrier of size 25cm x 45cm.
 The Vertical Autoclave's all joints should be TIG Welded with inert gas and smooth finished for Crevice free internals. The autoclave should have hydraulic die press seamless lid from stainless locking s/m steel plate. The autoclave lid should have locking / system for lid. Should have insulated material (industrial plastic) – cool to touch type. The lid cover should be provided with auto purge cum vacuum breaker valve and a manually operable valve for exhaust. The unit should have a solenoid valve for additionally purging the air & end of cycle automatic exhaust. The Vertical Autoclave should have stainless steel pressure gauge with dual range dial display in KPA and PSI along with a correlated temperature scale for steam in degrees Celsius. The user should be able to set the temperature up to 122°C in steps of 0.1°C each. The user should be able to set the Sterilization hold time in steps of 1 minutes each. The timer range should be upto 95 minutes In case of abrupt power failure / switching off, the Last Cycle which was set should remain in the memory. The unit should be equipped with Low Water Detection unit and should give Audio- Visual alarm in case of Low Water in the chamber and cut off the supply to the heater. The autoclave should give sensor open alarm in case sensor fault. The autoclave should be updio-visual alarm on completion of set autoclave cycle. The autoclave is pressure equipment & hence from safety point of view chamber should be CE(NB) certified. Autoclave should be use CE(NB) certified. Autoclave should be use DE(NB) certified. Autoclave should be SUB Single phase. The autoclave should be CE(NB) certified. Autoclave should be SUB Single phase. 	٠	
 plate. The autoclave lid should have locking / system for lid. Should have insulated material (industrial plastic) – cool to touch type. The lid cover should be provided with auto purge cum vacuum breaker valve and a manually operable valve for exhaust. The unit should have a solenoid valve for additionally purging the air & end of cycle automatic exhaust. The Vertical Autoclave should have stainless steel pressure gauge with dual range dial display in KPA and PSI along with a correlated temperature scale for steam in degrees Celsius. The vertical Autoclave should be controlled by a microprocessor-based controller. The user should be able to set the temperature up to 122°C in steps of 0.1°C each. The user should be able to set the Sterilization hold time in steps of 1 minutes each. The timer range should be upto 95 minutes In case of abrupt power failure / switching off, the Last Cycle which was set should remain in the memory. The display (LED/LCD) for the parameters should be two line Alpha memoric digital display. The unit should be equipped with Low Water Detection unit and should give Audio-Visual alarm in case of Low Water in the chamber and cut off the supply to the heater. The unit should have safety valve to protect the equipment in case of over pressurization. The unit should be ensured by inbuilt MCB. The autoclave chamber should be hydrostatically tested at 1.5 times of its working pressure. The Vertical Autoclave is pressure equipment & hence from safety point of view chamber should be designed as per ASME guidelines & should have PED certified. (From NB) The Vertical Autoclave should be CE(NB) certified. Autoclave should be provided with two wire mesh carriers of SS304. Supply Voltage: 230 V AC, 50 HZ, Single phase. The autoclave should be provided with Calibration reports with NABL traceability. 	•	The Vertical Autoclave's all joints should be TIG Welded with inert gas and smooth finished
 (industrial plastic) – cool to touch type. The lid cover should be provided with auto purge cum vacuum breaker valve and a manually operable valve for exhaust. The unit should have a solenoid valve for additionally purging the air & end of cycle automatic exhaust. The Vertical Autoclave should have stainless steel pressure gauge with dual range dial display in KPA and PSI along with a correlated temperature scale for steam in degrees Celsius. The operations of the unit should be controlled by a microprocessor-based controller. The user should be able to set the temperature up to 122°C in steps of 0.1°C each. The timer range should be upto 95 minutes In case of abrupt power failure / switching off, the Last Cycle which was set should remain in the memory. The display (LED/LCD) for the parameters should be two line Alpha memoric digital display. The unit should be equipped with Low Water Detection unit and should give Audio- Visual alarm in case of Low Water in the chamber and cut off the supply to the heater. The unit should give indication by audio-visual alarm on completion of set autoclave cycle. The electrical safety should be ensured by inbuilt MCB. The autoclave chamber should be hydrostatically tested at 1.5 times of its working pressure. The Vertical Autoclave should be CK(NB) certified. Autoclave should be reguiped with Calibration reports with NABL traceability. Manufacturer shall be ISO 13485 certified & should submit photocopy for the same. 	•	
 operable valve for exhaust. The unit should have a solenoid valve for additionally purging the air & end of cycle automatic exhaust. The Vertical Autoclave should have stainless steel pressure gauge with dual range dial display in KPA and PSI along with a correlated temperature scale for steam in degrees Celsius. The operations of the unit should be controlled by a microprocessor-based controller. The user should be able to set the temperature up to 122°C in steps of 0.1°C each. The timer range should be upto 95 minutes In case of abrupt power failure / switching off, the Last Cycle which was set should remain in the memory. The display (LED/LCD) for the parameters should be two line Alpha memoric digital display. The autoclave should give sensor open alarm in case sensor fault. The unit should have safety valve to protect the equipment in case of over pressurization. The unit should give indication by audio-visual alarm on completion of set autoclave cycle. The autoclave should be hydrostatically tested at 1.5 times of its working pressure. The Vertical Autoclave is pressure equipment & hence from safety point of view chamber should be designed as per ASME guidelines & should have PED certified. (From NB) The Vertical Autoclave should be CE(NB) certified. Autoclave should be supplied with two wire mesh carriers of SS304. Supply Voltage: 230 V AC, 50 HZ, Single phase. The autoclave should be provided with Calibration reports with NABL traceability. 	•	
 automatic exhaust. The Vertical Autoclave should have stainless steel pressure gauge with dual range dial display in KPA and PSI along with a correlated temperature scale for steam in degrees Celsius. The operations of the unit should be controlled by a microprocessor-based controller. The user should be able to set the temperature up to 122°C in steps of 0.1°C each. The user should be able to set the Sterilization hold time in steps of 1 minutes each. The timer range should be upto 95 minutes In case of abrupt power failure / switching off, the Last Cycle which was set should remain in the memory. The display (LED/LCD) for the parameters should be two line Alpha memoric digital display. The unit should be equipped with Low Water Detection unit and should give Audio- Visual alarm in case of Low Water in the chamber and cut off the supply to the heater. The unit should have safety valve to protect the equipment in case of over pressurization. The unit should give indication by audio-visual alarm on completion of set autoclave cycle. The autoclave chamber should be hydrostatically tested at 1.5 times of its working pressure. The Vertical Autoclave is pressure equipment & hence from safety point of view chamber should be designed as per ASME guidelines & should have PED certified. (From NB) The Vertical Autoclave should be CE(NB) certified. Autoclave should be supplied with two wire mesh carriers of SS304. Supply Voltage: 230 V AC, 50 HZ, Single phase. The autoclave should be provided with Calibration reports with NABL traceability. 	•	
 display in KPA and PSI along with a correlated temperature scale for steam in degrees Celsius. The operations of the unit should be controlled by a microprocessor-based controller. The user should be able to set the temperature up to 122°C in steps of 0.1°C each. The user should be able to set the Sterilization hold time in steps of 1 minutes each. The timer range should be upto 95 minutes In case of abrupt power failure / switching off, the Last Cycle which was set should remain in the memory. The display (LED/LCD) for the parameters should be two line Alpha memoric digital display. The unit should be equipped with Low Water Detection unit and should give Audio- Visual alarm in case of Low Water in the chamber and cut off the supply to the heater. The autoclave should give sensor open alarm in case sensor fault. The unit should have safety valve to protect the equipment in case of over pressurization. The electrical safety should be ensured by inbuilt MCB. The autoclave chamber should be hydrostatically tested at 1.5 times of its working pressure. The Vertical Autoclave is pressure equipment & hence from safety point of view chamber should be designed as per ASME guidelines & should have PED certified. (From NB) The Vertical Autoclave should be CE(NB) certified. Autoclave should be supplied with two wire mesh carriers of SS304. Supply Voltage: 230 V AC, 50 HZ, Single phase. The autoclave should be provided with Calibration reports with NABL traceability. 	•	
 The user should be able to set the temperature up to 122°C in steps of 0.1°C each. The user should be able to set the Sterilization hold time in steps of 1 minutes each. The timer range should be upto 95 minutes In case of abrupt power failure / switching off, the Last Cycle which was set should remain in the memory. The display (LED/LCD) for the parameters should be two line Alpha memoric digital display. The unit should be equipped with Low Water Detection unit and should give Audio- Visual alarm in case of Low Water in the chamber and cut off the supply to the heater. The autoclave should give sensor open alarm in case sensor fault. The unit should have safety valve to protect the equipment in case of over pressurization. The unit should give indication by audio-visual alarm on completion of set autoclave cycle. The autoclave chamber should be hydrostatically tested at 1.5 times of its working pressure. The Vertical Autoclave is pressure equipment & hence from safety point of view chamber should be designed as per ASME guidelines & should have PED certified. (From NB) The Vertical Autoclave should be CE(NB) certified. Autoclave should be supplied with two wire mesh carriers of SS304. Supply Voltage: 230 V AC, 50 HZ, Single phase. The autoclave should be provided with Calibration reports with NABL traceability. Manufacturer shall be ISO 13485 certified & should submit photocopy for the same. 	٠	display in KPA and PSI along with a correlated temperature scale for steam in degrees
 The user should be able to set the Sterilization hold time in steps of 1 minutes each. The timer range should be upto 95 minutes In case of abrupt power failure / switching off, the Last Cycle which was set should remain in the memory. The display (LED/LCD) for the parameters should be two line Alpha memoric digital display. The unit should be equipped with Low Water Detection unit and should give Audio- Visual alarm in case of Low Water in the chamber and cut off the supply to the heater. The autoclave should give sensor open alarm in case sensor fault. The unit should have safety valve to protect the equipment in case of over pressurization. The electrical safety should be ensured by inbuilt MCB. The autoclave chamber should be hydrostatically tested at 1.5 times of its working pressure. The Vertical Autoclave is pressure equipment & hence from safety point of view chamber should be designed as per ASME guidelines & should have PED certified. (From NB) The Vertical Autoclave should be CE(NB) certified. Autoclave should be supplied with two wire mesh carriers of SS304. Supply Voltage: 230 V AC, 50 HZ, Single phase. The autoclave should be provided with Calibration reports with NABL traceability. Manufacturer shall be ISO 13485 certified & should submit photocopy for the same. 	•	The operations of the unit should be controlled by a microprocessor-based controller.
 The timer range should be upto 95 minutes In case of abrupt power failure / switching off, the Last Cycle which was set should remain in the memory. The display (LED/LCD) for the parameters should be two line Alpha memoric digital display. The unit should be equipped with Low Water Detection unit and should give Audio- Visual alarm in case of Low Water in the chamber and cut off the supply to the heater. The autoclave should give sensor open alarm in case sensor fault. The unit should have safety valve to protect the equipment in case of over pressurization. The unit should give indication by audio-visual alarm on completion of set autoclave cycle. The autoclave chamber should be hydrostatically tested at 1.5 times of its working pressure. The Vertical Autoclave is pressure equipment & hence from safety point of view chamber should be designed as per ASME guidelines & should have PED certified. (From NB) The Vertical Autoclave should be CE(NB) certified. Autoclave should be supplied with two wire mesh carriers of SS304. Supply Voltage: 230 V AC, 50 HZ, Single phase. The autoclave should be provided with Calibration reports with NABL traceability. Manufacturer shall be ISO 13485 certified & should submit photocopy for the same. 	•	The user should be able to set the temperature up to 122°C in steps of 0.1°C each.
 The timer range should be upto 95 minutes In case of abrupt power failure / switching off, the Last Cycle which was set should remain in the memory. The display (LED/LCD) for the parameters should be two line Alpha memoric digital display. The unit should be equipped with Low Water Detection unit and should give Audio- Visual alarm in case of Low Water in the chamber and cut off the supply to the heater. The autoclave should give sensor open alarm in case sensor fault. The unit should have safety valve to protect the equipment in case of over pressurization. The unit should give indication by audio-visual alarm on completion of set autoclave cycle. The autoclave chamber should be hydrostatically tested at 1.5 times of its working pressure. The Vertical Autoclave is pressure equipment & hence from safety point of view chamber should be designed as per ASME guidelines & should have PED certified. (From NB) The Vertical Autoclave should be CE(NB) certified. Autoclave should be supplied with two wire mesh carriers of SS304. Supply Voltage: 230 V AC, 50 HZ, Single phase. The autoclave should be provided with Calibration reports with NABL traceability. Manufacturer shall be ISO 13485 certified & should submit photocopy for the same. 	•	The user should be able to set the Sterilization hold time in steps of 1 minutes each.
 in the memory. The display (LED/LCD) for the parameters should be two line Alpha memoric digital display. The unit should be equipped with Low Water Detection unit and should give Audio- Visual alarm in case of Low Water in the chamber and cut off the supply to the heater. The autoclave should give sensor open alarm in case sensor fault. The unit should have safety valve to protect the equipment in case of over pressurization. The unit should give indication by audio-visual alarm on completion of set autoclave cycle. The electrical safety should be ensured by inbuilt MCB. The autoclave chamber should be hydrostatically tested at 1.5 times of its working pressure. The Vertical Autoclave is pressure equipment & hence from safety point of view chamber should be designed as per ASME guidelines & should have PED certified. (From NB) The Vertical Autoclave should be CE(NB) certified. Autoclave should be supplied with two wire mesh carriers of SS304. Supply Voltage: 230 V AC, 50 HZ, Single phase. The autoclave should be provided with Calibration reports with NABL traceability. Manufacturer shall be ISO 13485 certified & should submit photocopy for the same. 	•	The timer range should be upto 95 minutes
 display. The unit should be equipped with Low Water Detection unit and should give Audio- Visual alarm in case of Low Water in the chamber and cut off the supply to the heater. The autoclave should give sensor open alarm in case sensor fault. The unit should have safety valve to protect the equipment in case of over pressurization. The unit should give indication by audio-visual alarm on completion of set autoclave cycle. The electrical safety should be ensured by inbuilt MCB. The autoclave chamber should be hydrostatically tested at 1.5 times of its working pressure. The Vertical Autoclave is pressure equipment & hence from safety point of view chamber should be designed as per ASME guidelines & should have PED certified. (From NB) The Vertical Autoclave should be CE(NB) certified. Autoclave should be supplied with two wire mesh carriers of SS304. Supply Voltage: 230 V AC, 50 HZ, Single phase. The autoclave should be provided with Calibration reports with NABL traceability. Manufacturer shall be ISO 13485 certified & should submit photocopy for the same. 	•	
 alarm in case of Low Water in the chamber and cut off the supply to the heater. The autoclave should give sensor open alarm in case sensor fault. The unit should have safety valve to protect the equipment in case of over pressurization. The unit should give indication by audio-visual alarm on completion of set autoclave cycle. The electrical safety should be ensured by inbuilt MCB. The autoclave chamber should be hydrostatically tested at 1.5 times of its working pressure. The Vertical Autoclave is pressure equipment & hence from safety point of view chamber should be designed as per ASME guidelines & should have PED certified. (From NB) The Vertical Autoclave should be CE(NB) certified. Autoclave should be supplied with two wire mesh carriers of SS304. Supply Voltage: 230 V AC, 50 HZ, Single phase. The autoclave should be ISO 13485 certified & should submit photocopy for the same. 	•	
 The unit should have safety valve to protect the equipment in case of over pressurization. The unit should give indication by audio-visual alarm on completion of set autoclave cycle. The electrical safety should be ensured by inbuilt MCB. The autoclave chamber should be hydrostatically tested at 1.5 times of its working pressure. The Vertical Autoclave is pressure equipment & hence from safety point of view chamber should be designed as per ASME guidelines & should have PED certified. (From NB) The Vertical Autoclave should be CE(NB) certified. Autoclave should be supplied with two wire mesh carriers of SS304. Supply Voltage: 230 V AC, 50 HZ, Single phase. The autoclave should be provided with Calibration reports with NABL traceability. Manufacturer shall be ISO 13485 certified & should submit photocopy for the same. 	•	
 The unit should give indication by audio-visual alarm on completion of set autoclave cycle. The electrical safety should be ensured by inbuilt MCB. The autoclave chamber should be hydrostatically tested at 1.5 times of its working pressure. The Vertical Autoclave is pressure equipment & hence from safety point of view chamber should be designed as per ASME guidelines & should have PED certified. (From NB) The Vertical Autoclave should be CE(NB) certified. Autoclave should be supplied with two wire mesh carriers of SS304. Supply Voltage: 230 V AC, 50 HZ, Single phase. The autoclave should be provided with Calibration reports with NABL traceability. Manufacturer shall be ISO 13485 certified & should submit photocopy for the same. 	•	The autoclave should give sensor open alarm in case sensor fault.
 The electrical safety should be ensured by inbuilt MCB. The autoclave chamber should be hydrostatically tested at 1.5 times of its working pressure. The Vertical Autoclave is pressure equipment & hence from safety point of view chamber should be designed as per ASME guidelines & should have PED certified. (From NB) The Vertical Autoclave should be CE(NB) certified. Autoclave should be supplied with two wire mesh carriers of SS304. Supply Voltage: 230 V AC, 50 HZ, Single phase. The autoclave should be provided with Calibration reports with NABL traceability. Manufacturer shall be ISO 13485 certified & should submit photocopy for the same. 	•	The unit should have safety valve to protect the equipment in case of over pressurization.
 The autoclave chamber should be hydrostatically tested at 1.5 times of its working pressure. The Vertical Autoclave is pressure equipment & hence from safety point of view chamber should be designed as per ASME guidelines & should have PED certified. (From NB) The Vertical Autoclave should be CE(NB) certified. Autoclave should be supplied with two wire mesh carriers of SS304. Supply Voltage: 230 V AC, 50 HZ, Single phase. The autoclave should be provided with Calibration reports with NABL traceability. Manufacturer shall be ISO 13485 certified & should submit photocopy for the same. 	•	The unit should give indication by audio-visual alarm on completion of set autoclave cycle.
 The Vertical Autoclave is pressure equipment & hence from safety point of view chamber should be designed as per ASME guidelines & should have PED certified. (From NB) The Vertical Autoclave should be CE(NB) certified. Autoclave should be supplied with two wire mesh carriers of SS304. Supply Voltage: 230 V AC, 50 HZ, Single phase. The autoclave should be provided with Calibration reports with NABL traceability. Manufacturer shall be ISO 13485 certified & should submit photocopy for the same. 	٠	The electrical safety should be ensured by inbuilt MCB.
 should be designed as per ASME guidelines & should have PED certified. (From NB) The Vertical Autoclave should be CE(NB) certified. Autoclave should be supplied with two wire mesh carriers of SS304. Supply Voltage: 230 V AC, 50 HZ, Single phase. The autoclave should be provided with Calibration reports with NABL traceability. Manufacturer shall be ISO 13485 certified & should submit photocopy for the same. 	•	The autoclave chamber should be hydrostatically tested at 1.5 times of its working pressure.
 Autoclave should be supplied with two wire mesh carriers of SS304. Supply Voltage: 230 V AC, 50 HZ, Single phase. The autoclave should be provided with Calibration reports with NABL traceability. Manufacturer shall be ISO 13485 certified & should submit photocopy for the same. 	•	
 Supply Voltage: 230 V AC, 50 HZ, Single phase. The autoclave should be provided with Calibration reports with NABL traceability. Manufacturer shall be ISO 13485 certified & should submit photocopy for the same. 	•	The Vertical Autoclave should be CE(NB) certified.
 The autoclave should be provided with Calibration reports with NABL traceability. Manufacturer shall be ISO 13485 certified & should submit photocopy for the same. 	•	Autoclave should be supplied with two wire mesh carriers of SS304.
• Manufacturer shall be ISO 13485 certified & should submit photocopy for the same.	•	Supply Voltage: 230 V AC, 50 HZ, Single phase.
· · · ·	•	The autoclave should be provided with Calibration reports with NABL traceability.
Local Service Setup for prompt and efficient post-sales support.	•	Manufacturer shall be ISO 13485 certified & should submit photocopy for the same.
	•	Local Service Setup for prompt and efficient post-sales support.

1 1

149

ŝ,

*

Consignee Details

	M/s. Equitron Medica Private Limited			
	Autoclave Machine			
	Make: Equitron			
	Model : #7410-4072			
Delivery Per	iod 06 weeks			
	No.: 7746 /Haffkine/Procurement Cell/E-4444/Autoc	No.: 7746 /Haffkine/Procurement Cell/E-4444/Autoclave Machine/		
PO Ref. No.	2022-23.			
	Date: 15/10/22			
	प्रशासकीय मान्यता निधी २४,००,०००/-			
	राज्य योजना २०२१-२२			
Sr. No.	Name & Address of the Consignee (1			
1.	CAMA & Albless Hospital, Mumbai.	04		
	Total :	04		

Dr. Sadanand Bhise Manager General Manager Haffkine Bio Pharmaceutical Corporation Ltd. (Procurement Cell), Mumbai.

मा. व्यवस्थापकीय संचालक यांच्या मान्यतेने व करित्व

, .

