

**HAFFKINE BIO PHARMACEUTICAL CORPORATION LIMITED****Procurement Cell****( A Government of Maharashtra Undertaking)****Regd. Office : Acharya Donde Marg, Parel, Mumbai 400 012 (INDIA)****Phone No: 022- 24129320-23**  
**Managing Director :022-24150628**  
**General Manager (Procurement**  
**Cell) :022-24100478****Website : http://www.vaccinehaffkine.com****E-mail: procurementcell@vaccinehaffkine.com****No.: 4640 /Haffkine/Procurement Cell/E-1340/ Monitor Vial**  
**Sign, NIBP, HR, SpO2, ECG, RR, Temp/ 2020-21.****Date: 18/05/2021**

To,  
**M/s. BPL Medical Technologies Pvt. Ltd.,**  
11<sup>th</sup> KM, Bannerghatta Road,  
Arekere, Bangalore -560076.  
Contact No.: 9867571651.  
**E-Mail: [sashi.bhushan@bpl.in](mailto:sashi.bhushan@bpl.in)**

**Subject : Supply Order for Tender No. E-1340/Monitor Vial Sign, NIBP, HR, SpO2, ECG, RR, Temp.**

**Reference: 1. Tender No. E-1340/HBPCL/PC/Monitor Vial Sign, NIBP, HR, SpO2, ECG, RR, Temp/ 2019-20.**

**2. शासननिर्णय, क्रमांक - प्रशामा-०११९/प्र.क्र.३४/आरोग्य-७**

**दिनांक : २८ जानेवारी, २०१९.**

**3. Sanction of Tender Approval Committee Meeting Dated 23.03.2021.**

With reference to the tender cited under reference no 1, you are requested to supply the following goods as per details mentioned below to consignee list enclosed with this order.

Sr. No.	Name of the item	Specification of item	Quantity/ Unit (DHS)	Unit Rate inclusive of GST(Rs.)	Total Amount Rs.
1	Monitor Vial Sign, NIBP, HR, SpO2, ECG, RR, Temp Make : BPL Model : Excello	As per Annexure X	07	1,20,960/-	8,46,720/-

**Total amount in words: Rupees Eight Lakhs Fourty Six Thousand Seven Hundred Twenty Only.**

**Factory Location: BPL Medical Technologies Pvt. Ltd., BPL Works Palakkad, Kerala – 678007.**

- Forwarding:** Forwarding Free on Road Destination. I.e. door delivery basis.
- Delivery Period:** 06 weeks from the date of receipt of order by the supplier to the consignee attached.
- Pre-Dispatch Inspection:** Supplier shall make necessary arrangement / facilitate to carry out Pre-Dispatch inspection as per Tender Terms & condition and submit the Inspection report to this office. The Pre-Dispatch inspection cost will be borne by supplier. Machine should be dispatched only after satisfactory Pre-Dispatch Inspection.

- 4 **Risk purchase clause:** If the bidder fails to supply the stores within the stipulated delivery period, the order will stand cancelled. Undersigned shall be entitled to purchase such stores from any other source at such price which ordinarily should not be more than 10% of the tender price. The extra expenditure in such cases shall be recovered by Managing Director, Haffkine Bio Pharmaceutical Corporation Ltd. (Procurement Cell), Mumbai from the Supplier.
- 5 **Payment Terms:** Payment of 100% of the contract value will be made within 8 weeks on delivery and successful installation and satisfactory commissioning and operation of the machinery.
- 6 **Acceptance & Receipt:** It should be submitted in Appropriate Format to the purchasing authority.
- 7 **Delivery Challan** - Should be sent in the name of consignee in duplicate. It should specify Name of Equipment / Mfg. by / packing & quantity.
- 8 **Invoice Copy** - Should be sent in triplicate on the Name of Managing Director, Haffkine Bio Pharmaceutical Corporation Ltd.(Procurement Cell), Mumbai. Along with Bill of Entry and Country of Origin Certificate of the consignment.
- 9 **Other Terms :**
  - 1) **Warranty:** The warranty period shall be for 2 years from the date of commissioning of all equipment supplied as certified by the consignee. After completion of 2 years warranty period Manufacturer/Supplier should give commitment to ensure services and supply of spare part for further 8 years. The successful tenderer must ensure 95% uptime during warranty period. In case of downtime, warranty period will be extended for period of downtime. If the equipment is not attended within 24 hours for Mumbai and 48 hours for other places the supplier will be liable to pay a penalty of 0.07% of purchase cost for every day of delay. Such penalty will be recovered from the amount of security deposit. Certificate of such uptime / downtime issued by the end user will be binding for the supplier Replacement of spares parts thereof due to manufacturing defects during warranty period will be entirely at the supplier's cost.
  - 2) The user institution will enter to the Comprehensive Maintenance Contract with supplier agency @ 5% of the order value (excluding taxes) of the equipment per year for 8 years after completion of warranty period. In case of non-compliance of CMC the supplier will be liable to pay penalty or for appropriate action. Payment of CMC on yearly basis will be made by the user's institution, at the end of the year after satisfactory performance report from the end user.
- 10 **Contract Agreement:** Bidder should submit Contract Agreement on non-judicial stamp paper of requisite value.

#### **Fall Clause**

It is a condition of the contract that all through the currency thereof, the price at which you will the supply stores should not exceed the lowest price charged by you to any customer during the currency of the contract and that in the event of the prices going down below the rate contract prices you shall promptly furnish such information to us to enable to amend the contract rates for subsequent supplies.

4.2	<b>Battery operated :</b> Battery charger to be integral to mains power supply, and to charge battery during mains power operation of unit. battery powered, silenceable alarm for power failure. Internal, replaceable, rechargeable battery allows operation for at least one hour in the event of power failure. Battery backup of minimum 100 minutes.
4.3	<b>Tolerance (to variations, shutdowns):</b> Voltage corrector/stabilizer to allow operation at $\pm 30\%$ of local rated voltage.
4.4	<b>Protection :</b> Electrical protection provided by fuses in both live and neutral supply lines.
4.5	<b>Power consumption:</b> <120Watt
4.6	<b>Other energy supplies :</b> Mains cable
5	<b>ACCESSORIES, SPARE PARTS, CONSUMABLES</b>
5.1	<b>Accessories &amp; Spares:</b> 2 pairs, 3 to 5 lead ECG cable 2 packs of 100 disposable ECG connection electrodes. Two sets of reusable SpO2 probes including neonatal probes. Two sets of NIBP cuffs of each size. Two external skin temperature probes. Expiry date of consumable should be 6 months.
5.2	Consumables / reagents (open, closed system)
6	<b>ENVIRONMENTAL AND DEPARTMENTAL CONSIDERATIONS</b>
6.1	<b>Atmosphere / Ambiance (air conditioning, humidity, dust....):</b> Capable of being stored continuously in ambient temperature of 0 to 50 deg C and relative humidity of 15 to 90% Capable of operating continuously in ambient temperature of 10 to 40 deg C and relative humidity of 15 to 90%
6.2	<b>User's care, cleaning, Disinfection &amp; Sterility issues:</b> Complete unit to be easily washable and sterilizable using alcohol and other chemical agents.
7	<b>STANDARDS AND SAFETY</b>
7.1	<b>Certifications:</b> The machine should be CE Certified (From Notified Body) or US FDA Approval and Company should be ISO 13485 Certified.
8	<b>TRAINING AND INSTALLATION</b>
8.1	<b>Pre-installation requirements:</b> nature, values, quality, tolerance : Supplier to perform installation, safety and operation checks before handover.
8.2	<b>Requirements for sign-off:</b> Supplier to perform installation, safety and operation checks before handover local clinical staff to affirm completion of installation.
8.3	<b>Training of staff (medical, paramedical ,technicians):</b> Training of users in operation and basic maintenance shall be provided
8.4	Others
9	<b>WARRANTY AND MAINTENANCE</b>
9.1	<b>Warranty:</b> 2 years
9.2	<b>Maintenance tasks:</b> Maintenance manual detailing complete maintaining schedule
9.3	<b>Service contract clauses, including prices:</b> Warranty of 2 years with free servicing (min.3/year) during warranty
10	<b>DOCUMENTATION</b>
10.1	<b>Operating manuals, service manuals, other manuals:</b> Advanced maintenance tasks required shall be documented User, technical and maintenance manuals to be supplied in English language. List to be provided of equipment and procedures required for local calibration and routine maintenance
10.2	<b>Other accompanying documents :</b> List to be provided of important spares and accessories, with their part numbers and cost. Certificate of calibration and inspection to be provided.

11	<b>NOTES</b>
11.1	<b>Other information:</b> Contact details of manufacture, supplier and local service agent to be provided
11.2	<b>Recommendations or warnings:</b> Any warning signs would be adequately displayed

**Consignee Details**

<b>M/s. BPL Medical Technologies Pvt Ltd.</b>		
<b>Monitor Vial Sign, NIBP, HR, SpO2, ECG, RR, Temp</b> <b>Make : BPL, Model : Excello</b>		
<b>Delivery Period</b>	06 weeks	
<b>PO Ref. No.</b>	No.: 4640 /Haffkine/Procurement Cell/E-1340/ Monitor Vial Sign, NIBP, HR, SpO2, ECG, RR, Temp/2020-21. Date: 18/05/2021	
<b>Sr. No.</b>	<b>Name &amp; Address of the Consignee</b>	<b>Qty.</b>
1)	SDH Manchar, Pune.	02
2)	SDH Umerga, Osmanabad.	02
3)	SDH, Pusad, Yavatmal.	01
4)	RH Akkalkuwa, Nandurbar.	02
	<b>Total</b>	<b>07</b>

*Bawiskar*  
*12/5*  
**Dr. Vijay P. Bawiskar**  
**General Manager - 1**

**Haffkine Bio Pharmaceutical Corporation Ltd.**  
**(Procurement Cell), Mumbai.**